



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER OF PATENTS AND TRADEMARKS  
Washington, D.C. 20231  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/687,606	10/11/2000	Enrique J. Klein	020460 000710US	2574

20350 7590 11/29/2001

TOWNSEND AND TOWNSEND AND CREW  
TWO EMBARCADERO CENTER  
EIGHTH FLOOR  
SAN FRANCISCO, CA 94111-3834

EXAMINER

STEWART, ALVIN J

ART UNIT	PAPER NUMBER
----------	--------------

3738

DATE MAILED: 11/29/2001

7

Please find below and/or attached an Office communication concerning this application or proceeding.

# Office Action Summary

Application No.  
09/687,606

Applicant(s)  
Klein

Examiner  
Alvin Stewart

Art Unit  
3738



-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

## Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136 (a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

1) ☒ Responsive to communication(s) filed on Oct 11, 2000

2a) ☐ This action is FINAL. 2b) ☒ This action is non-final.

3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11; 453 O.G. 213.

## Disposition of Claims

4) ☒ Claim(s) 1-5 is/are pending in the application.

4a) Of the above, claim(s) \_\_\_\_\_ is/are withdrawn from consideration.

5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.

6) ☒ Claim(s) 1-5 is/are rejected.

7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.

8) ☐ Claims \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

9) ☐ The specification is objected to by the Examiner.

10) ☒ The drawing(s) filed on Oct 11, 2000 is/are objected to by the Examiner.

11) ☐ The proposed drawing correction filed on \_\_\_\_\_ is: a) ☐ approved b) ☐ disapproved.

12) ☐ The oath or declaration is objected to by the Examiner.

## Priority under 35 U.S.C. § 119

13) ☐ Acknowledgement is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d).

a) ☐ All b) ☐ Some\* c) ☐ None of:

- ☐ Certified copies of the priority documents have been received.
- ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
- ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\*See the attached detailed Office action for a list of the certified copies not received.

14) ☒ Acknowledgement is made of a claim for domestic priority under 35 U.S.C. § 119(e).

## Attachment(s)

15) ☒ Notice of References Cited (PTO-892)

16) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)

17) ☒ Information Disclosure Statement(s) (PTO-1449) Paper No(s). 6

18) ☐ Interview Summary (PTO-413) Paper No(s). \_\_\_\_\_

19) ☐ Notice of Informal Patent Application (PTO-152)

20) ☐ Other:

Art Unit: 3738

## DETAILED ACTION

### *Drawings*

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the coating of radiopaque material over the stainless steel stent which varies in thickness over an axial length must be shown or the feature(s) canceled from the claim(s). No new matter should be entered.

### *Claim Rejections - 35 USC § 102*

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless --

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371© of this title before the invention thereof by the applicant for patent.

Claims 1, 4 and 5 are rejected under 35 U.S.C. 102(e) as being anticipated by Lashinski et al US Patent 6,071,296.

Lashinski et al discloses a stent made of radiopaque material wherein the radiopaque material is thicker near the ends of the cylindrical frame than over the midsection (see col. 5, lines 63-66 through col. 6, lines 1-5).

Art Unit: 3738

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2 and 3 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lashinski et al US Patent 6,071,296 in view of Callol et al US Patent 6,174,329.

Lashinski et al discloses the invention substantially as claimed. Additionally, Lashinski et al discloses a stent that can be altered and be changed with different material composition (e.g. metal or non-metal)(see col. 4, lines 64-67). However, Lashinski et al does not disclose a frame made of stainless steel and a coating made of gold, platinum, etc...

Callol et al teaches a stent having a frame made of stainless steel and a coating made of gold, tantalum, etc., so it can twist and deform easily and it can be seen from a fluoroscope.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to alter and change the material composition of the frame with the stainless steel core and gold coat of the Callol et al reference in order to change the material property of the stent so it can deform more easily and it can be seen from a fluoroscope.

Art Unit: 3738

*Conclusion*

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

US 5,607,442 See Abstract.

US 5,725,572 See Figs. 2A-5.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Exr. Alvin Stewart whose telephone number is (703) 305-0277. The examiner can normally be reached on Monday-Friday from 7:30 AM to 5:00 PM.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott, can be reached on (703)308-2111. The fax phone number for the organization where this application or proceeding is assigned is (703) 305-3590.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is (703) 308-0858.



Alvin Stewart

November 14, 2001.



CORRINE McDERMOTT  
SUPERVISORY PATENT EXAMINER  
TECHNOLOGY CENTER 3700